




BioCode[®]


Fungal Panel

Package Insert



Applied BioCode

Barcoded Magnetic Beads | Flexible Syndromic Tests



BioCode® Fungal Panel Package Insert

Part Number 64-F0001

For Research Use Only. Not for use in diagnostic procedures.

Contact Information

Customer Service:

Telephone: 1-562-777-9800

Email: Orders@apbiocode.com

Website:

www.apbiocode.com

Technical Services:

Telephone: 1-833-BMB-Tech

(1-833-262-8324)

Email: TechSupport@apbiocode.com

Mailing Address:

12130 Mora Drive, Unit 2

Santa Fe Springs, CA 90670, USA



Applied BioCode, Inc.

12130 Mora Drive, Unit 2

Santa Fe Springs, CA 90670, USA

For Research Use Only. Not for use in diagnostic procedures.

Table of Contents

INTRODUCTION	3
SUMMARY AND EXPLANATION OF THE TEST	3
PRINCIPLE OF PROCEDURE	3
Overview of a BioCode® MDx-3000 Run	3
MATERIALS REQUIRED	4
Materials Provided with Each Kit	4
Materials Required but Not Provided with Each Kit	5
WARNINGS AND PRECAUTIONS	6
General Precautions	6
Laboratory Precautions	6
Safety Precautions	6
REAGENT STORAGE, HANDLING, AND STABILITY	6
MDx-3000 FUNGAL PANEL PROTOCOL SETUP PROCEDURE	7
ASSAY PROCEDURE	13
Nucleic Acid Purification (for Example Purposes Only)	13
BioCode® FP Assay Setup	13
INTERPRETATION OF RESULTS	14
External Control	14
Internal Control	15
Target Pathogen Interpretation	15
BioCode® Fungal Panel Reports	15
TABLE OF SYMBOLS	17

INTRODUCTION

BioCode® Fungal Panel

The BioCode® Fungal Panel (FP) is a multiplex nucleic acid amplification test intended for use with the BioCode® MDx-3000 Instrument. The BioCode® FP is a qualitative assay capable of the simultaneous detection and identification of nucleic acids from the fungal pathogens (mold, dimorphic fungi, and yeast) listed in Table 1.

Table 1. Organisms detected by the BioCode® Fungal Panel

Mold	
<ul style="list-style-type: none"> ▪ <i>Aspergillus fumigatus</i> ▪ <i>Aspergillus terreus</i> ▪ <i>Aspergillus flavus</i> ▪ <i>Aspergillus niger</i> ▪ <i>Mucor indicus</i> ▪ <i>Mucor circinelloides/racemosus</i> ▪ <i>Rhizopus</i> ▪ <i>Rhizomucor</i> 	<ul style="list-style-type: none"> ▪ <i>Cunninghamella bertholletiae</i> ▪ <i>Lichtheimia corymbifera^a</i> ▪ <i>Scedosporium apiospermum</i> ▪ <i>Lomentospora prolificans</i> ▪ <i>Fusarium oxysporum</i> ▪ <i>Fusarium solani</i> ▪ <i>Syncephalastrum racemosum^b</i> ▪ <i>Pneumocystis jirovecii</i>
Dimorphic Fungi	Yeast
<ul style="list-style-type: none"> ▪ <i>Blastomyces dermatitidis</i> ▪ <i>Coccidioides immitis/posadasii</i> ▪ <i>Histoplasma capsulatum</i> 	<ul style="list-style-type: none"> ▪ <i>Cryptococcus neoformans/gattii</i>

^aAssay may also detect some strains of other *Lichtheimia* species, such as *L. ramosa*.

^bAssay may also detect some strains of other *Syncephalastrum* species, such as *S. monosporum*.

SUMMARY AND EXPLANATION OF THE TEST

The BioCode® FP is a multiplex nucleic acid test intended for the qualitative detection of deoxyribonucleic acids (DNA) from 20 fungal pathogens. The BioCode® FP targets the highly conserved ITS2 region between the 5.8S and 28S rDNA for all fungi while *Pneumocystis jirovecii* is detected by its large mitochondrial subunit. Additionally, the assay targets a portion of the 18S fungal rDNA to detect the order Mucorales including *Rhizopus*, *Mucor*, and *Rhizomucor*. The BioCode® FP is designed for use with the BioCode® MDx-3000 automated system, and test results are available in about 4 hours post-DNA extraction. The test is for research use only and not for use in diagnostic procedures.

PRINCIPLE OF PROCEDURE

The BioCode® MDx-3000 is an automated system that integrates PCR amplification, target capture, signal generation, and optical detection of nucleic acids for multiple fungal pathogens.

Overview of a BioCode® MDx-3000 Run

- Multiplex PCR** – Extracted nucleic acid is added to a freshly prepared multiplex PCR reaction mix to amplify the target nucleic acids present in the sample. In each primer pair, one primer is biotinylated at the 5'-end to produce labeled PCR product for subsequent detection.
- Dispensing BMB-Probe Mix** – As PCR amplification nears completion, the instrument's robotic head dispenses BMB-Probe Mix into the assigned reaction wells of the capture plate using disposable pipette tips.

For Research Use Only. Not for use in diagnostic procedures.

3. **PCR Product Transfer** – Following the completion of PCR amplification, the robotic head punctures the foil seal with disposable pipette tips to acquire the PCR amplicons and transfers them into the appropriate wells of the capture plate.
4. **Target Capture** – Biotin-labeled PCR amplicons are captured at a specific temperature by target-specific probes linked to designated Barcoded Magnetic Beads (BMBs). Throughout this process, the BMBs remain suspended through gentle agitation. To differentiate captured targets, each pathogen and the internal control is assigned a unique BMB pattern.
5. **Signal Generation** – Following the removal of both unbound PCR amplicons and primers via washing, the robot dispenses streptavidin-phycoerythrin (SA-PE) conjugate into the reaction. The strong affinity between biotin and SA-PE ensures that PCR amplicons captured with the biotin moiety are labeled with phycoerythrin in close proximity to the BMBs.
6. **Optical Detection** – Optical detection is conducted for each reaction well of the optically clear, 96-well flat-bottom microtiter capture plate. Following the removal of unbound SA-PE conjugates through washing, excitation of the fluorophore at the designated wavelength elicits a fluorescence signal from BMBs tagged with SA-PE conjugates. Each reaction well is imaged at a specific emission wavelength for fluorescent signal and under bright field for identifying the barcode patterns (decoding).
7. **Software Analysis** – The BioCode® MDx-3000 Software controls the operation of the instrument, collects and analyzes data, and automatically generates interpretation for test reports at the end of the run. Fluorescent signals from BMBs with the same barcode are sorted and calculated to generate median fluorescence intensity (MFI) for each analyte. For the BioCode® FP, the presence or absence of a pathogen can be determined relative to the MFI thresholds specified by the user. The software also analyzes the results of optional external and internal controls to validate the run and individual specimen results.

MATERIALS REQUIRED

Materials Provided with Each Kit

Table 2. Reagents provided in the BioCode® FP Assay Kit – enough for 96 tests

Component Name	Part No.	Contents	Storage
BioCode® Master Mix A	14-M0001	500 µL x 2	Store at -20°C before use. After thaw, store at 4°C for up to 30 days.
BioCode® Fungal Primer Mix	24-F0006	500 µL x 2	Store at -20°C before use. After thaw, store at 4°C for up to 30 days.
BioCode® DNA-IC	23-D0001	500 µL x 2	Store at -20°C before use. After thaw, store at 4°C for up to 30 days.
BioCode® Fungal BMB-Probe Mix	24-F0007	6000 µL x 1	Store at -20°C before use. After thaw, vortex for 30 sec, store at 4°C up to 90 days.

Materials Required but Not Provided with Each Kit

Table 3. General Reagents required for the BioCode® MDx-3000

Component Name	Part No.	Contents	Storage
BioCode® SA-PE Mix	63-S0001	450 µL x 8	Single use; protect from light; store at 4°C. Do Not Freeze
BioCode® Buffer A	44-B0003	1 L x 1	Store at room temperature

Table 4. BioCode® MDx-3000 Consumables

Reagent	Source/Part No.	Quantity
Reagent Reservoirs	Applied BioCode 01-R0005 or INTEGRA 4332	50 each x 4
Waste Bin and Lid	Applied BioCode 01-W0104 and 01-W0105	25 each
20 µL pipette tips	Applied BioCode 01-P0006 or Beckman 717256	10 x 96 tips
250 µL pipette tips	Applied BioCode 01-P0007 or Beckman 717252	10 x 96 tips
Bio-Rad 96-well hard shell plate 0.1 mL	Applied BioCode 01-P0011 or Bio-Rad HSL9601	25 PCR plates
PCR Adhesive Foil	Applied BioCode 01-P0012 or Thermo Fisher Scientific AB-0626 or Eppendorf 0030127790	100 foils
Microtiter plate	Applied BioCode 01-P0009 or Greiner bio-one 655101	10 plates or 10 plates x 10 bags
Microtiter plate lid	Applied BioCode 01-P0010 or Nunc 5500	50 lids or 100 lids

Equipment and Additional Consumables Required (but Not Provided)

- BioCode® MDx-3000
- Nucleic acid extraction method (e.g., MagNA Pure 96 [Roche], EMAG [bioMérieux])
- Vortex
- Centrifuge
- Pipettes – single, multi-channel and/or repeater with accuracy range between 1-10 µL, 10-200 µL, and 100-1000 µL
- Sterile, RNase/DNase-free disposable aerosol-barrier micro pipettor tips
- 1.5 mL polypropylene micro centrifuge tubes and racks (RNase/DNase free recommended)
- Cooler racks for 1.5 mL tubes and 0.1 mL 96 well plate
- Freezer (manual defrost) at -10 to -30°C
- Freezer (manual defrost) at -60 to -90°C
- Refrigerator at 2 to 8°C

WARNINGS AND PRECAUTIONS

General Precautions

1. For Research Use Only. Not for use in diagnostic procedures.
2. The BioCode® FP Assay is to be used with the BioCode® MDx-3000.
3. Contamination may occur if carry over of samples is not adequately controlled during sample handling and processing or from the environment.

Laboratory Precautions

1. The BioCode® FP Assay should be performed in clearly defined work areas moving in one direction from pre-amplification areas to the amplification/detection area to reduce potential for contamination.
 - a. Begin with reagent preparation before moving to amplification/detection.
 - b. Use dedicated equipment and supplies for each area (including personal protective equipment, such as lab coats and disposable gloves).
 - c. Clean work areas with 10% bleach or similar disinfectant followed by 70% ethanol before and after assay preparation.
2. A negative control is recommended for each run. If multiple lots are assayed at the same time, a negative control is recommended for each lot.
3. Do not use reagents past the expiration date. It is not recommended to mix reagents or interchange kit components from different kit lots. Kit configurations are identified on the kit outer box.
4. Assay setup should be performed at room temperature.

Safety Precautions

1. Follow universal safety procedures.
2. Dispose of unused kit reagents according to local, state, and federal regulations.
3. Wear appropriate personal protective equipment including, but not limited to, lab coats, gloves, and protective eyewear. Change gloves often.
4. Safety Data Sheets for all reagents provided in the BioCode® FP can be accessed at <https://www.apbiocode.com/resources>.
5. Do not pipette by mouth.

REAGENT STORAGE, HANDLING, AND STABILITY


1. Store the kit components frozen (-30°C to -10°C) prior to use.
2. Once thawed, store Master Mix, Primer Mix, and DNA-IC refrigerated (2-8°C) for up to 30 days.
3. Once thawed, store BMB-Probe Mix refrigerated (2-8°C) for up to 90 days. Do not combine leftover BMB-Probe Mix from different kits or lots.
4. SA-PE mix is for single use only. Store refrigerated (2-8°C). Protect from light. **DO NOT FREEZE.**
5. Store Buffer A at room temperature (15-25°C).
6. Avoid storage of any materials near heating or cooling vents or in direct sunlight.
7. Always check the expiration date and do not use reagents beyond the expiration date printed.
8. Remove BMB-Probe Mix from MDx-3000 once the run is complete and store refrigerated (2-8°C).

MDx-3000 FUNGAL PANEL PROTOCOL SETUP PROCEDURE


The BioCode® FP assay is designed to run with the User-Defined Mode (UDM) of the MDx-3000 software. The user is responsible for creating and saving their own UDM threshold detection protocol as well as determining their own assay threshold values and Minimum BMB Count.

Refer to the BioCode® MDx-3000 User Manual for more detailed instructions on how to set up a UDM run.


1. On the MDx-3000 software main screen, select User-Defined Mode.




2. On the next screen, select Manage UDM Protocol.




3. On the next screen, select New.




4. Enter a name for the Fungal Panel protocol and select Next.




- For “Step 1 of 4: Thermal Cycler”, toggle the switch on the top right of the screen from Off to On and fill in the PCR cycling parameters as shown below. Select Next when finished.




- For “Step 2 of 4: BMB Dispense and Target Capture”, fill in the BMB Dispense, Transfer From Thermal Cycler, and Hybridization and Washing parameters as shown below. Select Next when finished.



- For “Step 3 of 4: SAPE Dispense and Incubation”, fill in the Fluorescent Dye Dispense and Incubation and Washing parameters as shown below. Select Next when finished.



- For “Step 4 of 4: Optical Detection”, toggle the switch on the top right of the screen from Off to On, then toggle the Threshold Detection switch from Off to On and fill in Background Subtraction Barcode Number, Internal Control Barcode (IC), and Barcode Description as shown below. Enter a Barcode Threshold value between 500 and 99999 for the Internal Control Barcode. For this Barcode Threshold, the screenshot provided below is only an example – the user is responsible for determining their own assay threshold value for the Internal Control. Enter a Minimum BMB Count value between 6 and 100. For Minimum BMB Count, the screenshot provided below is only an example – the user is responsible for determining their own Minimum BMB Count.



- On the left side of the screen, enter the Barcode and Description information from Table 5 below into the Optical Detection table.

Table 5. BioCode® FP Barcode numbers, descriptions, and detected targets.

Barcode ^a	Description	Target Detected
24	Lich spp	<i>Lichtheimia</i> spp.
25	Sync spp	<i>Syncephalastrum</i> spp.
26	M circ/race	<i>Mucor circinelloides/racemosus</i>
27	Mucorales	Mucorales
32	A fum	<i>Aspergillus fumigatus</i>
33	A terr	<i>Aspergillus terreus</i>
34	A flav	<i>Aspergillus flavus</i>
35	A nig	<i>Aspergillus niger</i>
36	C bert	<i>Cunninghamella bertholletiae</i>
37	M ind	<i>Mucor indicus</i>
38	Rhizomuc spp	<i>Rhizomucor</i> spp.
39	Rhizopus spp	<i>Rhizopus</i> spp.
42	P jiro	<i>Pneumocystis jirovecii</i>
43	F oxy	<i>Fusarium oxysporum</i>
44	F sol	<i>Fusarium solani</i>
47	S apio	<i>Scedosporium apiospermum</i>
48	L pro	<i>Lomentospora prolificans</i>
59	Crypto spp	<i>Cryptococcus</i> spp.
60	B derm	<i>Blastomyces dermatitidis</i>
61	C imm/posa	<i>Coccidioides immitis/posadasii</i>
62	H cap	<i>Histoplasma capsulatum</i>

^aBarcode/target thresholds to be determined by user


- Enter threshold values for each barcode into the Optical Detection table. The screenshots provided below are only an example. The user is responsible for determining their own assay threshold values for each barcode in the panel.

Optical Detection

All Barcodes
 Defined Barcodes

Barcode	Description	Threshold
24	Lich spp	750
25	Sync spp	750
26	M circ/race	750
27	Mucorales	750
32	A fum	750
33	A terr	750
Barcode	Description	Threshold
34	A flav	750
35	A nig	750
36	C bert	750
37	M ind	750
38	Rhizomuc spp	750
39	Rhizopus spp	750
Barcode	Description	Threshold
42	P jiro	750
43	F oxy	750
44	F sol	750
47	S apio	750
48	L pro	750
59	Crypto spp	750
60	B dermat	750
61	C imm/posa	750
62	H cap	750

11. When finished with Step 4 of 4, click Save. Then, click Complete.



ASSAY PROCEDURE

The starting material for the BioCode® FP assay is purified nucleic acid. The user should determine the method of nucleic acid purification used and should ensure that method is compatible with PCR. The BioCode® DNA-IC provided in the kit should be added to each sample prior to extraction and is used as an extraction, amplification, and detection control for the BioCode® FP.

A negative control is recommended for each run. If multiple BioCode® FP kit lots are assayed at the same time, a negative control is recommended for each kit lot. A negative control may consist of extracted negative matrix or a well characterized negative sample spiked with BioCode® DNA-IC.

Gloves and other personal protective equipment (PPE) should be used when handling reagents. Once PCR reagents are prepared and sample is added to the PCR plate, the plate should be promptly transferred to the instrument to start the run. After the run is complete, the PCR plate and capture plate should be sealed and discarded.

Refer to the BioCode® MDx-3000 User Manual for more detail and pictorial representations of the BioCode® MDx-3000 User-Defined Mode run set up instructions, as well as instructions for viewing UDM result reports.

Nucleic Acid Purification (for Example Purposes Only)

Note: The method used for nucleic acid purification is determined by the user. This section is provided only as an example.

1. Prepare samples of interest by combining 200 µL sample, 400 µL 1x PBS buffer, and 10 µL DNA-IC into a bead beating tube.
2. Vortex the tube with beads on high for 10 minutes. Briefly spin down the tube in a centrifuge.
3. Extract nucleic acids with a method selected by the user.

BioCode® FP Assay Setup

Note: Prepare the PCR Plate in a dedicated reaction mix preparation area.

1. Thaw Primer Mix, Master Mix A, and BMB-Probe Mix at room temperature. For the Primer Mix and

For Research Use Only. Not for use in diagnostic procedures.

Master Mix A, perform a quick vortex (2-3 seconds) and centrifuge to collect reagents at the bottom of the tubes.

2. Prepare the reaction mix in a polypropylene microcentrifuge tube as described below:

Table 6. Reaction Mix Formulation

Component	Reaction Mix Volume (µL) per reaction	Reaction Mix Volume (µL) per 10 reactions
BioCode® Master Mix A	10.0 µL	100.0 µL
BioCode® Fungal Panel Primer Mix	10.0 µL	100.0 µL
Reaction Mix Volume (µL)	20 µL	200 µL

3. Mix reaction mix by briefly pulse vortexing or pipetting up and down 8 to 10 times. Centrifuge to collect contents at the bottom of the tube. If proceeding immediately to PCR plate setup, reaction mix may be kept at room temperature. If not proceeding immediately to PCR plate setup, store reaction mix at 2-8°C or in a cooling block for up to 1 hour.
4. Pipette 20 µL of reaction mix into the appropriate wells of a 96-well plate.
5. Pipette 5 µL of each extracted sample into the wells.
6. If using extracted negative control(s), pipette 5 µL of the negative control(s) into the negative control well(s).
7. Seal plate with a pierceable foil. Store at 2-8°C or on a cooling block until ready to load onto the BioCode® MDx-3000 (not to exceed one hour from the time the reaction mix is prepared).
8. Briefly centrifuge plate to collect samples at the bottom of the plate.
9. Load plate onto the BioCode® MDx-3000.
10. Vortex BMB-Probe Mix for 30 seconds at high speed and load the vial onto the BioCode® MDx-3000. **(Note: Precipitates may appear at cold temperatures. If precipitates are present, allow the BMB-Probe Mix to warm to room temperature and vortex additional 30 seconds.)** Based on the number of wells required in the plate setup, the BioCode® MDx-3000 software calculates the bulk BMB-Probe Mix volume required and displays the volume on the screen. The BioCode® MDx-3000 automatically dispenses 50 µL of BMB-Probe Mix into each well of the optical detection plate, based on the plate setup.
11. Load reagents and consumables as prompted by graphic user interface. Refer to the BioCode® MDx-3000 User Manual for more details on UDM run setup. Start the Run.


INTERPRETATION OF RESULTS

The BioCode® MDx-3000 software will analyze data based on plate validity, sample validity, and Median Fluorescent Intensity (MFI) compared to the MFI thresholds specified by the user during the BioCode® FP UDM threshold detection protocol setup.

External Control

An external negative control is recommended and can be negative sample matrix or well characterized negative samples spiked with the provided DNA-IC. If the user decides to include a

negative control in a BioCode® FP run, it should go through all processing steps (extraction, amplification, and detection). During run setup, an external negative control can be defined by naming any sample “Negative Control” (case insensitive) in the Sample ID column on the left side of the “Input Sample IDs” screen, as exemplified in the screenshot below.



Internal Control

It is recommended that the user add the DNA Internal Control (DNA-IC) supplied in the kit to each sample and negative control during extraction. The internal control monitors the efficiency of the extraction, amplification, and detection stages of the assay. Lack of DNA-IC signal may indicate sample-associated inhibition or reagent/instrument issues. Samples suspected of being inhibitory can be repeated from extraction. If reagent or instrument issues are suspected, samples may be repeated from stored nucleic acid extracts.

Target Pathogen Interpretation

Fluorescent signals from BMBs with the same barcode are sorted and the median fluorescence intensity (MFI) is calculated for each analyte. The BioCode® MDx software interprets the results based on the MFI threshold values set by the user during the initial Fungal Panel UDM threshold detection protocol setup.

If the user decides to include controls in a run, such as DNA-IC in extracted samples and one or more external controls, it is recommended that the user examine the results of the control(s) prior to assessing the results for all other samples.

BioCode® Fungal Panel Reports

The analyzed BioCode® MDx-3000 results are displayed in four report formats: MFI Report, BMB Count Report, UDM Run Report, and Sample Reports. All reports can be reviewed electronically and exported as a PDF or CSV file. The MFI Report and BMB Count Report display raw results for specimens and controls, while the UDM Run Report and Sample Reports contain interpreted results. Refer to the BioCode® MDx-3000 User Manual for more details and example screenshots of BioCode® MDx-3000 UDM reports.

The MFI Report displays the raw MFI values calculated for each well. The user may compare the raw MFIs in the MFI Report to the barcode-specific MFI thresholds selected during the initial BioCode® MDx-3000 UDM

protocol setup.

The BMB Count Report displays the BMB counts for each barcode in each well. The user may compare the raw counts in the BMB Count Report to the Minimum Bead Count specified during the initial Fungal Panel UDM threshold detection protocol setup.


The UDM Run Report displays interpreted results in a tabular format for each well. Possible results by target are Detected, Not detected, or Invalid. The MFI threshold for Detected vs. Not detected is determined by the barcode thresholds specified during the initial Fungal Panel UDM threshold detection protocol setup.

The UDM Sample Reports display results for a single well. In addition to results for each target, the UDM Sample Reports include a results summary section which allows positive results to be reviewed at a glance. The UDM Sample Reports indicate well validity based on BMB counts, background MFI, and external and internal controls. UDM Sample Reports also include any sample-specific comments entered during run setup.

All report headers provide traceability information for the unique Run Name, UDM Protocol used, Run Start and Run Complete times, User, Run Status, SA-PE Mix lot (if entered by user during run setup), Wash buffer lot (if entered by user during run setup), Software Version/Mode, Instrument, and External Controls. The Run Status section will specify if the run is Valid, Invalid, Complete, or Incomplete based on the negative control results (if using one or more negative controls) for the specific run/kit lot. The External Controls section indicates the results for the negative controls (Valid, Invalid, or N/A). It is recommended to review the Run Status and External Controls sections prior to reviewing target results.

TABLE OF SYMBOLS

The following symbols are used on the BioCode® FP Assay components and/or in this package insert.

	Batch code		Contains 96 tests		Temperature limitations
	Part Number		Warning		Consult instructions for use
	Use by YYYY-MM-DD		Registered trademark		Manufacturer

The information contained in this document is subject to change without notice. No part of this document may be reproduced or transmitted in any form or by any means electronic or mechanical, for any purpose, without express written permission of Applied BioCode, Inc.

BioCode is a registered trademark of Applied BioCode, Inc. All other names of products and brands appearing in this manual are trademarks or registered trademarks of their respective owners.